

RANDOX

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2023-10-29
KIT DESCRIPTION:	Lipid Control (LPD CONTROL 3)
CATALOGUE NO:	LE2670
BATCH NO:	3102CH
EXPIRY DATE:	2026-10-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Level III / 3102CH	HDL- Cholesterol	2.06	1.75-2.37	1.89	mmol/l
Level III / 3102CH	LDL- Cholesterol	5.37	4.56-6.18	5.38	mmol/l

APPROVED BY:

M. Kelly

S

Verified by signNow
20-Jun-2024 16:03:14 UTC
805f292bb94c09921e
Mailing: Cobco

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S F Fitzgerald BSc PhD DSc (Hons) RAeMun. RJ Fitzgerald. W Fitzgerald OBE. R F Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB353030400.